



Declaration of Conformity

for sports & fitness devices

Manufacturer: h/p/cosmos sports & medical gmbh
Am Sportplatz 8
DE 83365 Nussdorf-Traunstein / Germany
phone +49 / 86 69 / 86 42 0 fax +49 / 86 69 / 86 42 49
email@h-p-cosmos.com www.h-p-cosmos.com

Product: Treadmill Ergometer (running machines)

Classification according to DIN EN 957-1	S = (Studio) Professional and / or commercial application A = high accuracy
UMDNS Code	-.-.-.-.-
GMDN Code	-.-.-.-.-
NBOG Code	-.-.-.-.-

Models:	h/p/cosmos stratos lt	h/p/cosmos stratos
	h/p/cosmos mercury lt	h/p/cosmos mercury
	h/p/cosmos stellar lt	h/p/cosmos stellar
	h/p/cosmos quasar lt	h/p/cosmos quasar

We herewith declare that the above mentioned types meet the following standards:

European Council Directive 2004/108 EEC (electromagnetic compatibility).

European Council Directive 2006/42 EEC (machinery directive)

European Council Directive 2006/95 EEC (low voltage directive)

The appendix **I** of the Directive 2004/108 EEC is applicable.

The appendix **I+II** of the Directive 2006/42 EEC is applicable.

The **CE**-mark gets affixed to the products.

The following listed and fulfilled Norms are elements of our production:

[DIN EN 60335-1](#), [DIN EN 61000-6-1](#), [DIN EN 61000-6-3](#),
[DIN EN 957-1](#), [DIN EN 957-6](#),
[DIN EN ISO 9001](#), [DIN EN ISO 14971](#)

All norms and standards listed in this document refer to validity date (year/month) as is was standard at the time/date when this document was printed and when the product has been manufactured. In case there was a transitional period of 2 valid norm editions at that time, then please ask h/p/cosmos or refer to the details as stipulated in the technical master file of the product for the precisely validity/issue date of the norm/standard.

This certificate of compliance covers all running machine models which are listed above and which are produced by h/p/cosmos on and from January 18, 2011. The validity of this certificate of compliance expires on the release date of a certificate of compliance with a later issue date, which may be necessary due to technical modifications or due to changes of standards, guidelines and directives.

DE 83365 Nussdorf-Traunstein, January 18, 2011

Franz Harrer
President & CEO

Wolfgang Petry
Safety Representative (Medical Devices § 30)